

34 1/2

7 C C & L

W. E. Dubois

Assayed

January 12/1880

The Standard
Weights no longer
reliable & should
be discarded & new
one sent from
Bureau of
Weights & Measures.

Recd. Jan. 12th 1880

Note to the Superintendent.

In reference to the letter from the Director, to ascertain the deviations of our set of standard weights, it appears upon further examination that they are no longer fit to occupy that position, but should be discarded, or used for common work; and a new set asked for, from the Bureau of Weights and Measures.

They have been on hand a long time, and considerably used; so much so, that they are somewhat battered, ~~with~~^{and} even small bits knocked off, and the loss restored by inserting pieces of lead. It would seem to be lost labor to make a serial correction, in weights intended to occupy a governing place, and so much out of order.

With respects of
W E Du Bois
assayer

[Abstract:] The Standard Weights no longer reliable & should be discarded... 34 ½

76602

Note to the Superintendent.

In reference to the letter from the Director, to ascertain the deviations of our set of standard weights, it appears upon further examination that they are no longer fit to occupy that position, but should be discarded, or used for common work; and a new set asked for, from the Bureau of Weights and Measures.

They have been on hand a long time, and considerably used; so much so, that they are somewhat battered, and even small bits knocked off, and the loss restored by inserting pieces of lead. It would seem to be lost labor to make a serial connection, in weights intended to occupy a governing place, and so much out of order.

With respects of
W.E. Du Bois
Assayer